

Aidite zirconia block User Instruction (3D Pro zir Ultra_GC)

[Before use]

Aidite zirconia blocks are produced by CIP technology and pre-sintered in low temperature. Although the product has some strength, because of porosity, please handle carefully. When you receive the product, please check as below. If there is something exceptional, please contact the sales manager or call +86-335-8587898.

- 1.Product is complete without any damage.
- 2.The packing is complete without any damage.
- 3.Label: company name, product name, batch number, inspectors and inspection date.

[Application Range]

Anterior full crown, Anterior full crown bridge, Posterior full crown, Posterior full crown bridge, Screw-retained bridge, Full arch long bridge, Inlay/ Onlay, Veneer, Anterior/ Posterior cutback crown, Anterior/ Posterior cutback crown bridge, Anterior/ Posterior coping, Anterior/ Posterior coping bridge

Note: The length of the above crowns and bridges should be within three units, including implant crowns and bridges.

[Color]

A1, A2, A3, A3.5, A4, B1, B2, B3, B4, C1, C2, C3, C4, D2, D3, D4, OM1, OM2, OM3

[Requirements for preparation]

Anterior crown:

The inner edge of shoulder must be slick or be fluted.

Preparation width of incisal, labial and palatal should be above 0.7mm.

The axial wall must be blunt to ensure the zirconia can reach an ideal milling effect with CAD/CAM.

Posterior crown:

The inner edge of shoulder must be slick or be fluted.

The minimum width of incisal edge, labial side, and palatal side is at least 0.7 mm.

Width of occlusal surface of lateral side of the lip, tongue preparation is more than 1.0 mm.

The axial wall must be blunt to ensure the zirconia can reach an ideal milling effect with CAD/CAM.

[Application method]




[Application method-scanning and designing]

Scanning

Select a high-precision scanner for scanning to obtain accurate restoration model data.

Precautions:

- Before scanning, check the impression and model for air bubbles, gypsum tumors, and layers; check the undercut of the abutment; check whether the abutment, adjacent teeth, and contralateral teeth are complete; check whether the occlusal space is suitable.
 - Check to make sure that the abutment fits perfectly with the base of the model.
- Regularly calibrate the scanner to ensure scanning accuracy.




Designing

It is designed according to the actual situation of the patient and the requirements of the doctor.

As an all-ceramic restoration, the following points need to be met:

- Please design according to the following minimum thickness requirements. The thinnest should be no less than 0.7mm.
- The cross-sectional area of the anterior connector should be at least 9mm² and the cross-sectional area of the posterior connector should be at least 12mm², connector height ≥3.
- No free pontic



[Application method-Nesting and Milling]

Nesting


Choose zirconia block of suitable thickness. When layout, put data in the middle of disc (vertical direction). The most suitable distance between tooth data and disc surface should be at least 0.5mm.

Connector should be placed in the most protruding position on buccal side.

The thickness of connector should be 2mm(as picture shows)

Place the disc:

Directing arrow is seen on the side of disc. The arrow points at the direction of incisal, namely the part of disc showing shallower color.



Milling

Before milling, check to see if the bur used is sharp enough and ensure the stability of the cutting equipment.

During processing, liquid cooling of zirconia blanks is not allowed. After milling

Is finished , check that if there are any defects happening listed below:

- Is there any crack?
- Is there any contamination?
- Is there any break?
- If any of these defects happen , reasons must be found and restorations need to be milled again.



[Application method-separating and cleaning]

Separating

Using technician specialized hand piece and grinding head to separate restorations from blocks.

Before grinding, a towel should be put on the desk to avoid restorations dropping on desk and crack or break. And then the connector need to be polished successively in one direction. Don't separate one connector Completely off at one time. Finally, the rest of connector could be polished slightly.

It is not advised to make too much adjustment on the restorations in soft condition to avoid causing sub fissure or chipping and soon .

Cleaning

Cleaning up the powder on surface and inner side of restorations with brush. If cleaning is not thorough , uncleaned powder will contaminate color liquid when dyeing and the uncleaned powder will stay on the surface and inner side of restorations after high temperature sintering, forming white spots and therefore having negative effect on esthetics and positioning of restorations.


[Application method-sintering]

--Before sintering, the furnace temperature must be calibrated, and the restorations must be dried.

--Compatible with CSF-400 fast sintering furnace, it enables ultra-fast sintering of restorations with 3 units or less. Note that two layers of crucibles must be placed for each sintering, and the upper layer must be covered with a lid to ensure effectiveness and temperature stability.

--Compatible with CSF-400Pro sintering furnace, it can sinter full arch long bridge.

--For full arch long bridges, a sintering rack must be added, and the “standing-up” placement method is recommended; for short units, the restoration with lingual side or occlusion facing down should be put on the beads in sagger.




Sintering program

The following curve is applicable to all sintering furnaces. If you want customized curves for different sintering furnaces or faster sintering curves, please contact us.

Below 6 units bridge (4h)

No.	Rate (°C/min)	Temp (°C)	Holding time (min)
1	30*	1050	0
2	5	1550	60
3	20	800	0

No.	Temp/°C	Time/min
1	20	34
2	1050	100
3	1550	60
4	1550	37
5	800	-121




*Note: Different sintering furnaces have different heating rates, and the first stage heating rate can be adjusted according to the furnace performance

Above 7 units bridge(11.5h)

No.	Rate (°C/min)	Temp (°C)	Holding time (min)
1	5	900	30
2	3	1530	120
3	8	100	0


No.	Temp/°C	Time/min
1	20	176
2	900	30
3	900	210
4	1530	120
5	1530	154
6	100	-121



- Note:
- 1.The above curve is only for reference, as there may be temperature differences in each sintering furnace without calibration. It is recommended to place a temperature measuring ring during the sintering process and observe whether there are any abnormalities in the dental effect. If the sintering effect is abnormal, combined with the temperature measurement results of the temperature measuring ring, we can provide guidance for adjusting the curve.
 - 2.When using the long bridge sintering curve, the furnace opening temperature should not exceed 100 °C

[Zirconium beads]

- When the zirconium beads are severely discolored, the shape is broken or damaged, it must be replaced immediately.
- If the zirconium beads are stuck together, be sure to break them apart to ensure proper bead function.
- The amount of zirconium beads should completely cover the bottom of the box (2 - 3layers).
- When replacing zirconium beads, first sinter the zirconium beads with remnants of green state zirconia and conduct a normal sintering cycle.
- It is recommended that Aidite zirconium Beads be used and it is recommended to use zirconium beads with a diameter less than or equal to 1.0mm to sinter long bridges. Use zirconium beads with a diameter greater than 1.2mm to sinter single crown.



[sintering furnace]

- The sintering furnace must use a voltage regulator to ensure stable operating voltage.
- The sintering furnace must be cleaned regularly (once a week) and kept dry. Cleaning method: scrape of the impurities in the furnace.
- Place green-state scrap zirconia scraps into the furnace and sinter them according to the normal zirconia sintering curve.
- If furnace has not been used for more than a week, it must be decontaminated before used.
- When the equipment is not in use, the furnace should be closed to ensure a dry environment inside the furnace. Please keep the operation room of the sintering equipment clean and free of dust and debris. Do not place sintering furnace in a dusty environment. Metal shavings or dust, can adversely affect the heating elements.
- The heating elements of the sintering furnace must not show damage. If there is a small amount of peeling on the surface of the heating rod (silicon-molybdenum rod) , the leftover material can be burned and the sinteringfurnace will back to normal.
- Check the furnace temperature regularly (every 3 months) to ensure the stability of the furnace temperature.
- Be sure to sinter in strict accordance with Aidite standard curve.
- If there is a power outage or shutdown of the sintering furnace during the sintering process, it is possible to observe whether the repair body crystallizes (shrinks). If crystallization has not occurred, it can be re sintered. If crystallization has already occurred, it needs to be re cut and made.
- If the sintering furnace temperature is not accurate, it may lead to poor sintering effect. The specific situation is as follows:
 1. Low furnace temperature may cause low transparency, deep color, and poor glossiness of the repair body, and it is necessary to calibrate the furnace temperature;
 2. If the furnace temperature is too high, it may lead to light color, increased or decreased transparency (material burning), and the furnace temperature needs to be calibrated.

[Application method-Grinding]

Use special zirconia grinding head to trim the surface of restoration.



Rough grinding:
This is the first grinding step after sintering restoration.The aim is to seat the restoration, adjust adjoining, occlusion and trim the anatomic contour of teeth.



Fine grinding:
Please make the tooth surface even and uniformly as well as the surface texture smoother.



Rough polishing:
To make the surface even and smooth.

Attention:

- 1.The cervical edge cannot be polished with rougher tools such as rough grinding or diamond grinding tools, otherwise it will increase the risk of fracture and even lead to the fracture ofthe restoration.
- 2.Only fine grinding and rough polishing tools can be used to polish the cervical edge to avoid chipping.
- 3.Do not grind the dental crown too roughly. After rough grinding, go over it again with fine grinding and rough polishing to make the surface smooth. If only rough grinding is carried out, it will result in dull glazing and the appearance of white spots on the surface.

[Application method-staining and Glazing]

Clean the surface of restoration by steam or ultrasonic cleaning machine before staining.
Biomic stain &glaze kit can help achieve better esthetics and effects.
For the specific operation method , please refer to the instru
Biomic stain &glaze kit.

